

Docket #: S21-245

Infrared Spectroscopy of Carbon Dioxide Hydration

Measurement of dissolved CO₂ has critical applications in healthcare monitoring and consumer goods quality control, yet is difficult to measure directly. Common methods include titration, measuring off-gas pressure, electrical conductivity or calculating chemical equilibria, all of which require a secondary calculating to determine the concentration of dissolved CO₂. Here researchers in the Dai lab have developed an infrared (IR) spectroscopy technique that allows them to overcome the typical challenge of CO₂ peaks being overshadowed by water. This ~100 micron IR cell creates a thin film of solution where the water does not absorb all the light, allowing the CO₂ signal to be resolved and quantified. This configuration tolerates high pressure systems (up to 58 atm) and still allows for clear CO₂ quantification. From CO₂ capture to consumer drinks and healthcare monitoring via dissolved CO₂ in blood, this technique provides a reliable, accurate and direct method for in situ measurement of dissolved CO₂.

Stage of Research

- Prototype

Applications

- Quantitative analysis of dissolved CO₂
- Consumer drinks: quality control and changing taste profiles of carbonated drinks
- Healthcare: measuring CO₂ levels in blood

Advantages

- Direct, in situ measurement of dissolved CO₂ vs measuring the gaseous phase and solving equilibria equations
- Capable of measuring high pressure or time-dependent systems

Publications

- Li, Jiachen, Jinyu Guo, and Hongjie Dai. "[Probing dissolved CO₂ \(aq\) in aqueous solutions for CO₂ electroreduction and storage.](#)" *Science advances* 8.19 (2022): eabo0399.

Patents

- Published Application: [WO2023069453](#)

Innovators

- Jiachen Li
- Hongjie Dai
- Jinyu Guo

Licensing Contact

Evan Elder

Associate Director, Licensing and Strategic Alliances, Physica

[Email](#)